



NOTES/SPECIFICATIONS:

1. CLEAR APERTURE: >90% DIAMETER
2. SURFACE ACCURACY:  $\lambda/4$  AT 632.8nm OVER CLEAR APERTURE
3. SURFACE QUALITY: 20-10 SCRATCH-DIG
4. BACK SURFACE: FINE GROUND
5. THICKNESS TOLERANCE:  $\pm 0.2$  mm
6. DIAMETER TOLERANCE: + 0.0/-0.2 mm
7. COATING: BROADBAND HR, DIELECTRIC  
 $R_s$  and  $R_p$  >99% OVER WAVELENGTH RANGE  
 AOI:  $0^\circ$  TO  $45^\circ$

FOR INFORMATION ONLY  
 NOT FOR MANUFACTURING PURPOSES

DRAWING PROJECTION			 www.thorlabs.com	
	NAME	DATE		
DRAWN	MG	15/JUN/11	$\varnothing 50.8\text{mm}$ , $f = 200.0\text{mm}$ BROADBAND MIRROR 350-400nm	
APPROVAL	PM	13/JUL/11	MATERIAL	REV A
COPYRIGHT © 2010 BY THORLABS			N-BK7	
VALUES IN PARENTHESIS ARE CALCULATED AND MAY CONTAIN ROUND OFF ERRORS			ITEM #	APPROX WEIGHT
			CM508-200-E01	0.05 kg