



NOTES/SPECIFICATIONS:

1. CLEAR APERTURE: >90% DIAMETER
2. SURFACE ACCURACY:  $\lambda/4$  AT 632.8nm OVER CLEAR APERTURE
3. SURFACE QUALITY: 20-10 SCRATCH-DIG
4. BACK SURFACE: FINE GROUND
5. THICKNESS TOLERANCE:  $\pm 0.2$  mm
6. DIAMETER TOLERANCE:  $+ 0.0/-0.2$  mm
7. COATING: BROADBAND HR, DIELECTRIC  
 $R_s$  and  $R_p$  >99% OVER WAVELENGTH RANGE  
 AOI:  $0^\circ$  TO  $45^\circ$

FOR INFORMATION ONLY  
 NOT FOR MANUFACTURING PURPOSES

DRAWING PROJECTION			 <a href="http://www.thorlabs.com">www.thorlabs.com</a>	
NAME	DATE			
DRAWN	MG	15/JUN/11	$\phi$ 75.0mm, f= 75.0mm BROADBAND MIRROR 350-400nm	
APPROVAL	PM	13/JUL/11	MATERIAL	REV
COPYRIGHT © 2010 BY THORLABS			N-BK7	A
VALUES IN PARENTHESIS ARE CALCULATED AND MAY CONTAIN ROUND OFF ERRORS			ITEM #	APPROX WEIGHT
			CM750-075-E01	0.19 kg