



NOTES/SPECIFICATIONS:

1. CLEAR APERTURE:  $> \phi 22.86\text{mm}$
2. SURFACE IRREGULARITY (S1):  $\leq \lambda/4$  AT 633nm OVER CLEAR APERTURE
3. BACK SURFACE (S2): UNIFORM FINE GRIND #12-#20 GRIT
4. SURFACE QUALITY(S1): 20-10 SCRATCH-DIG
5. THICKNESS TOLERANCE:  $\pm 0.1\text{ mm}$
6. DIAMETER TOLERANCE:  $+0.0/-0.2\text{ mm}$
7. DESIGN WAVELENGTH RANGE: 750-1100nm
8. COATING(S1): BROADBAND HR, DIELECTRIC  
 $R_s$  and  $R_p > 99\%$  AVERAGE OVER WAVELENGTH RANGE  
 AOI:  $0^\circ$  TO  $45^\circ$

FOR INFORMATION ONLY  
 NOT FOR MANUFACTURING PURPOSES

DRAWING PROJECTION			 www.thorlabs.com	
NAME	DATE			
DRAWN	YB	16/APR/18	$\phi 25.4\text{mm}$ , $f=150\text{mm}$ BROADBAND MIRROR 750-1100nm	
APPROVAL	DD	16/APR/18	MATERIAL	REV
COPYRIGHT © 2018 BY THORLABS			N-BK7	A
VALUES IN PARENTHESIS ARE CALCULATED AND MAY CONTAIN ROUND OFF ERRORS			ITEM #	APPROX WEIGHT
			CM254-150-E03	7.80 g